

# CALIBRATION CERTIFICATE

**Manufactured and calibrated by:** *Polimaster Ltd.*  
**Location of Calibration:** *51 F. Skoriny St., Minsk, 220141, Belarus*  
**Calibration Date:** *20.03.2020*  
**Item:** *Electronic Dosimeter PM1211-01*  
**Serial Number:** *192065*

**Environmental conditions**  
**Temperature:** *19,3 – 19,8 °C*  
**Relative Humidity:** *54,4 – 55,1 %*  
**Atmospheric Pressure:** *99,5 – 99,6 kPa*

DER	Background, μSv/h	0,8 μSv/h	800,0 μSv/h	80,0 mSv/h
Tolerance, %	–	±10,63	±10,04	±14,00
H <sub>j</sub>	0,08	0,89	785,49	81,01
Q <sub>j</sub>	–	1,3	-1,82	1,26
δ, %	–	4,61	4,83	4,61

The instrument is within the specified tolerance and passed the quality control.

Head of Quality Control Department



Department of Export Sales +375 17 396 3677,  
Research & Development Department +375 17 266 0527,  
Procurement Department +375 17 268 6757,  
Manufacturing Service +375 17 268 6819



Quality management system  
ISO 9001

- Customer focus
- Customer satisfaction
- Continuous improvement
- System/process effectiveness

ID 15 100 148764

www.tuv-thueringen.de