



Polimaster Ltd.

Head Office:

112, M. Bogdanovich str., Minsk, 220040

Phone: + 375 17 217 7080

Fax: + 375 17 217 7081

Manufacturing:

51, F. Skorina str., Minsk, 220141

Republic of Belarus

Phone: + 375 17 268 6819

Fax: + 375 17 260 2356

www.polimaster.com

E-mail: polimaster@polimaster.com



CALIBRATION CERTIFICATE

校正証明書
サーベイメータ PM1405

測定器のシリアル番号:

線量率	背景放射線量 μSv/h	0,8 μSv/h	30 μSv/h	80000 μSv/h
Tolerance, %	-			
平均線量率				
相対誤差[%]	-			
信頼限界	-			

Conformity Value, $\varphi_{oj}, \text{min}^{-1} \cdot \text{cm}^{-2}$	IUT Average Value, $\varphi_j, \text{min}^{-1} \cdot \text{cm}^{-2}$	Error Limit, $\delta, \%$	Allowable Error, $\delta_a, \%$
α-線 radiation			
β-線 radiation			

※校正方法の具体的な計算方法・手順については、取扱説明書またはWebサイトに記載しております。

The instrument is within the tolerance range and has passed the quality control.



Head of Quality Control Department:

Date of calibration: